



Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 9264  
 CALIBRATION DATE: 07-Mar-23

Slocum Payload CTD TEMPERATURE CALIBRATION DATA  
 ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = -1.497191e-04  
 a1 = 3.151397e-04  
 a2 = -5.055526e-06  
 a3 = 2.166138e-07

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	578611.6	1.0000	-0.0000
4.5000	495055.8	4.5001	0.0001
15.0000	316212.2	15.0000	-0.0000
18.5000	274027.0	18.5000	0.0000
24.0000	220128.0	23.9998	-0.0002
29.0000	181487.2	29.0003	0.0003
32.5000	159085.2	32.4998	-0.0002

n = Instrument Output (counts)

$$\text{Temperature ITS-90 (°C)} = 1 / \{ a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)] \} - 273.15$$

Residual (°C) = instrument temperature - bath temperature

