



Sea-Bird Scientific
 13431 NE 20th Street
 Bellevue, WA 98005
 USA

+1 425-643-9866
 seabird@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 9324
 CALIBRATION DATE: 24-Feb-21

Slocum Payload CTD PRESSURE CALIBRATION DATA
 1450 psia S/N 4417960

COEFFICIENTS:

PA0 =	3.046814e-02	PTCA0 =	5.254379e+05
PA1 =	4.638933e-03	PTCA1 =	-6.878708e+00
PA2 =	-2.918458e-11	PTCA2 =	3.053297e-01
PTEMPA0 =	1.599078e+02	PTCB0 =	2.537750e+01
PTEMPA1 =	-6.000413e-02	PTCB1 =	-1.000000e-04
PTEMPA2 =	-1.006424e-06	PTCB2 =	0.000000e+00

PRESSURE SPAN CALIBRATION

THERMAL CORRECTION

PRESSURE (PSIA)	INSTRUMENT OUTPUT (counts)	THERMISTOR OUTPUT (volts)	COMPUTED PRESSURE (PSIA)	RESIDUAL (%FSR)	TEMP (°C)	THERMISTOR OUTPUT (volts)	INSTRUMENT OUTPUT (counts)
14.79	528637.0	2193.0	14.84	0.00	32.50	2053	528729.20
315.10	593383.0	2189.0	315.08	-0.00	29.00	2107	528680.40
615.01	658116.0	2189.0	615.02	0.00	24.00	2185	528635.40
914.96	722909.0	2188.0	914.99	0.00	18.50	2270	528585.60
1214.96	787756.0	2185.0	1214.96	0.00	15.00	2324	528621.00
1464.93	841826.0	2188.0	1464.91	-0.00	4.50	2486	528581.60
1214.98	787760.0	2190.0	1214.99	0.00	1.00	2540	528630.60
915.04	722922.0	2187.0	915.05	0.00			
615.02	658116.0	2189.0	615.02	0.00	TEMPERATURE (°C)	SPAN	
315.08	593368.0	2197.0	315.03	-0.00	-5.00	25.38	
14.79	528623.0	2195.0	14.78	-0.00	35.00	25.37	

y = thermistor output (counts)

$$t = PTEMPA0 + PTEMPA1 * y + PTEMPA2 * y^2$$

$$x = \text{instrument output} - PTCA0 - PTCA1 * t - PTCA2 * t^2$$

$$n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)$$

$$\text{pressure (PSIA)} = PA0 + PA1 * n + PA2 * n^2$$

$$\text{Residual (\%FSR)} = (\text{computed pressure} - \text{true pressure}) * 100 / \text{Full Scale Range}$$

Date, Offset (%FSR)

● 24-Feb-21 -0.00

