PO Box 518
620 Applegate St.
Philomath, OR 97370

SEA-BIRD SCIENTIFIC

(541) 929-5650
Fax (541) 929-5277
www.Seabird.com

Scattering Meter Calibration Sheet

4/30/2026
Wavelength: 700 S/N FLBBCDSLC-4729

Use the following equation to obtain "scaled" output values:

$$\beta(\theta_c)\ m^{-1}\ sr^{-1} = \text{Scale Factor} \times (\text{Output} - \text{Dark Counts})$$

• **Scale Factor for 700 nm**	=	1.804E-06	$(m^{-1}sr^{-1})$/counts	
• **Output**	=	meter reading	counts	
• **Dark Counts**	=	48	counts	
Instrument Resolution	=	1.3	counts	2.32E-06 $(m^{-1}sr^{-1})$

Definitions:

- **Scale Factor**: Calibration scale factor, $\beta(\theta_c)$/counts. Refer to User's Guide for derivation.
- **Output**: Measured signal output of the scattering meter.
- **Dark Counts**: Signal obtained by covering detector with black tape and submersing sensor in water.

Instrument Resolution: Standard deviation of 1 minute of collected data.